

# Notice of Allowability

Application No.

10/649,350

Examiner

Nikita Wells

Applicant(s)

BOUSSE ET AL.

Art Unit

2881

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to Application filed 26 August 2003.
2. ☒ The allowed claim(s) is/are 1-12.
3. ☐ The drawings filed on \_\_\_\_\_ are accepted by the Examiner.
4. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☒ All b) ☐ Some\* c) ☐ None of the:
    1. ☒ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

## Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO-1449 or PTO/SB/08),  
Paper No./Mail Date 051004
4. ☐ Examiner's Comment Regarding Requirement for Deposit  
of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☐ Interview Summary (PTO-413),  
Paper No./Mail Date \_\_\_\_\_
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other \_\_\_\_\_

Nikita Wells  
Primary Examiner  
Art Unit: 2881

## Detailed Action

### *Allowable Subject Matter*

1. Claims 1-12 are allowed.
2. The following is an examiner's statement of reasons for allowance:

With respect to the independent claims 1, 5, and 7, prior art fails to disclose or make obvious, in addition to the accompanying features of the claims, a microfluidic chip and method for mass spectrometric analysis comprising: a microfluidic body formed with a plurality of reservoirs for containing at least one fluid medium; a plurality of separated channels formed within the microfluidic body that are in fluid communication with a corresponding reservoir; and an electrospray ionization tip formed with an open-ended distal tip portion that is positioned along an end surface of the microfluidic body, wherein the plurality of separated channels converge at the open-ended distal tip portion to direct an ionization spray derived from the fluid mediums.

The dependent claims 2-4, 6, and 8-12, are allowable by virtue of their dependence upon the independent claims 1, 5, and 7, respectively.

### *Drawings*

3. New corrected formal drawings are required in this application because the submitted drawings contain hand written labels, notation, and figure captions. Applicant is advised to employ the services of a competent patent draftsman outside the Office, as the U.S. Patent and Trademark Office no longer prepares new drawings. The corrected drawings are required in reply to the Office action to avoid abandonment of the application. The requirement for corrected drawings will not be held in abeyance.


***Conclusion***

4. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. A.T. Timperman (2003/0089605 A1) discloses a microfluidic system and method to analyze large numbers of compounds or complex mixtures of compounds using a plurality of sample holding channels which converge at an output channel. Moini et al. (6,465,776 B1) disclose a mass spectrometric apparatus for analyzing multiple fluid samples concurrently, having a plurality of separated channels converging, but fail to disclose that the convergence of these channels ends up at the open-ended distal tip portion of an ionization spray.
5. Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."
6. Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).
7. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Nikita Wells whose telephone number is (571) 272-2484. The examiner can normally be reached on 8:30 AM - 5:00 PM. If attempts to reach the examiner by

Art Unit: 2881

telephone are unsuccessful, the examiner's supervisor, John R. Lee can be reached on (571) 272-2477. The central fax phone number for the organization where this application or proceeding is assigned is (703) 872-9306.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is (703) 308-0956.

A handwritten signature in black ink, appearing to read "Nikita Wells". The signature is fluid and cursive, with the first name "Nikita" and the last name "Wells" clearly distinguishable.

Nikita Wells, Primary Examiner  
Art Unit 2881  
May 10, 2004